SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

ELLAB INC.
303 E 17th Ave, Suite D
Denver, CO 80203
Amanda Page Phone: 303-425-3370

CALIBRATION

Valid To: November 30, 2028 Certificate Number: 4828.02

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with R205 – A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations[1,2,3]:

I. Temperature

Parameter/Equipment	Range	CMC (±)	Comments
Temperature – Measuring Equipment	(-196 to -80) °C	0.012 °C (12 mK)	Liquid Nitrogen (-196 °C)
	(-80 to <0) °C	0.010 °C (10 mK)	Fluke 1595A
	0 °C	0.007 °C (7 mK)	SPRT 5699
	(>0° to 150) °C	0.011 °C (11 mK)	Ice Point
	(150 to 420) °C	0.015 °C (15 mK)	

[1] This laboratory offers commercial calibration service.

[2] Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

5202 Presidents Court, Suite 220 | Frederick, MD 21703-8398 | Phone: 301 644 3248 | Fax: 240 454 9449 | www.A2LA.org

[3] The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure the values in the ranges indicated for the listed measurement parameter.

Accredited Laboratory

A2LA has accredited

ELLAB INC.

Denver, CO

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).

Presented this 3rd day of November 2026.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 4828.02
Valid to November 30, 2028

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.